
Standard Method of Test for

**Determination of the Strength
of Soil–Lime Mixtures**

AASHTO Designation: T 220-66 (2018)

**Technical Subcommittee: 1a, Soil and Unbound Recycled
Materials**

Release: Group 3 (July)



**American Association of State Highway and Transportation Officials
444 North Capitol Street N.W., Suite 249
Washington, D.C. 20001**

Standard Method of Test for

Determination of the Strength of Soil–Lime Mixtures

AASHTO Designation: T 220-66 (2018)



Technical Subcommittee: 1a, Soil and Unbound
Recycled Materials

Release: Group 3 (July)

1. SCOPE

- 1.1. This method provides for the determination of the unconfined compressive strength of soil–lime stabilization mixtures.
- Note 1**—Generally, an unconfined compressive strength of 690 kPa (100 psi) is satisfactory for the final course of base construction, and it is desirable that materials for such courses contain a minimum of 50 percent plus 0.425-mm (No. 40) material before treatment. Various soil materials may be treated for subbase, and, in such cases, the minimum suggested unconfined compressive strength is 345 kPa (50 psi).
- 1.2. The values stated in SI units are to be regarded as the standard.

2. REFERENCED DOCUMENTS

- 2.1. *AASHTO Standards:*
- M 216, Lime for Soil Stabilization
 - R 74, Wet Preparation of Disturbed Soil Samples for Test
 - T 88, Particle Size Analysis of Soils
 - T 89, Determining the Liquid Limit of Soils
 - T 90, Determining the Plastic Limit and Plasticity Index of Soils
- 2.2. *ASTM Standards:*
- D2216, Standard Test Methods for Laboratory Determination of Water (Moisture) Content of Soil and Rock by Mass
 - E4, Standard Practices for Force Verification of Testing Machines
 - E11, Standard Specification for Woven Wire Test Sieve Cloth and Test Sieves

3. APPARATUS

- 3.1. *Automatic Tamper*—A compaction device with base plate to hold 152 mm (6 in.) ID forming molds, equipped with 4.54-kg (10-lb) ram and adjustable height of fall. Ram fall is 457 mm (18 in.). The striking face of the ram is a 40-degree segment of a circle of 76-mm (3-in.) radius. Automatic tamper shall be furnished an extra base plate to hold the forming mold during specimen top finishing.

- 3.2. *Compaction Mold with Removable Collar*—Mold is 152 mm (6 in.) ID and 215 mm (8½ in.) high.
- 3.3. *Measuring Device for Specimen Height*—A micrometer dial assembly with a standard set of spacer blocks.
- 3.4. *Scale*, rated 18.1-kg (40-lb) capacity, sensitive to 0.0005 kg (0.001 lb).
- 3.5. *Press*, hydraulic, to extrude specimens from mold.
- 3.6. *Drying Oven*, controlled from 60 ± 5°C (140 ± 9°F) to 110 ± 5°C (230 ± 9°F).
- 3.7. *A Supply of Metal Pans*, wide and shallow, for mixing and drying materials, and a supply of rectangular stainless steel pans approximately 230 by 400 by 60 mm (9 by 16 by 2¼ in.) deep, equipped with porous spacer plates.
- 3.8. *Circular Porous Stones*, slightly less than 152 mm (6 in.) in diameter and 51 mm (2 in.) high.
- 3.9. *Axial Cells*, lightweight stainless steel cylinders, 171 mm (6¾ in.) ID and 305 mm (12 in.) high, fitted with standard air valve and straight, tubular rubber membrane 152 mm (6 in.) in diameter.
- 3.10. *Vacuum Pump*, 20 to 35 L/min (0.70 to 1.25 cfm), or aspirator.
- 3.11. *Air Compressor*, 4.7 to 7.1 L/s (10 to 15 cfm) with 230-L (60-gal) storage tank and controls, pressure regulators, gauges, and valves.
- 3.12. *Moist Room*, equipped with shelves and a constant pressure supply of air.
- 3.13. *Micrometer Dial Gauge*, calibrated to 0.02 mm (0.001 in.) with support to measure deflection of specimen.
- 3.14. *A supply of 2.27- and 4.54-kg (5- and 10-lb) Lead Surcharge Weights*.
- 3.15. *Calibrated Proving Ring or Other Continuous Force Measuring Device*, according to ASTM E4, except that an error of plus or minus 2 percent is allowed.
- 3.16. *Circumference Measuring Device*—A specially made metal tape measure.
- 3.17. *Screw Jack Press and Assembly* or other suitable testing press, with dial housing and two loading blocks.
- 3.18. *Sieves* with square openings of the following sizes conforming to ASTM E11: 75-, 53-, 50-, 45-, 31.5-, 22.4-, 16.0-, 12.5-, 9.5-, 4.75-, 2.00-, 0.850-, and 0.425-mm (3.0-, 2.12-, 2.00-, 1.75-, 1.25-, 7/8-, 5/8-, 1/2-, 3/8-in., No. 4, No. 10, No. 20, and No. 40) sieves.
- 3.19. *Mechanical Pulverizer* with adjustable clearance rotating plate.
- 3.20. *Mechanical Sieve Shaker*—A laboratory size [0.014 m³ (½ ft³)] sieve shaker, convenient but not absolutely necessary, used for separating material for recombining specimens.
- 3.21. *A Supply of Small Tools and Accessories* such as wedgewood mortar and pestle, rubber-covered pestle, rawhide hammer, spatulas, trowels, level, scoops, siphon tubes, sample containers, cardboard cartons, 510-by-510-mm (20-by-20-in.) filter papers, etc.